



CERTIFICATE OF ACCREDITATION

ANSI-ASQ National Accreditation Board

500 Montgomery Street, Suite 625, Alexandria, VA 22314, 877-344-3044

This is to certify that

Hardy Diagnostics
1430 W. McCoy Lane
Santa Maria, CA 93455

has been assessed by ANAB
and meets the requirements of international standard

ISO/IEC 17025:2005

while demonstrating technical competence in the field of

CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of calibrations to which this accreditation applies.

AC-2546
Certificate Number


ANAB Approval

Certificate Valid: 01/16/2018-01/16/2020
Version No. 001 Issued: 01/16/2018



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Hardy Diagnostics
1430 W. McCoy Lane
Santa Maria, CA 93455
Alistair Wood 805-361-2695
wooda@hardydiagnostics.com

CALIBRATION

Valid to: **January 16, 2020**

Certificate Number: **AC-2546**

Mass

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Air Samplers	(100, 200) L/min	4 L/min	Wind Tunnel, Reference Anemometer, Barometer, Hygrometer

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2546.



Vice President

